

Verification of Compliance

Product Name : **Medical Type Switching Power Supply**
Model Number : **NMP1K2-aaaaaa-xx(a=C,E,H,K or # ; x=0~9)**
C=NMS-240-05, E=NMS-240-12,H=NMS-240-24,
K=NMS-240-48, #=Blank
Applicant : **MEAN WELL ENTERPRISES CO., LTD.**
Address : **No.28, Wuquan 3rd Rd., Wugu Dist., New Taipei City 248, Taiwan**
(R.O.C.)
Report Number : **S1C-M020-1706-361**
Issue Date : **October 2, 2018**
Applicable Standards :

Emission :

EN 60601-1-2 :2015 :
Home Healthcare Environment
EN 55011 : 2016+A1 : 2017 Group 1, Class B
EN 55032: 2015+AC:2016-07 Class B
EN 61000-3-2 : 2014
EN 61000-3-3 : 2013

Immunity :

EN 55024 : 2010+A1 :2015
EN 60601-1-2 :2015 : Home Healthcare Environment
EN 61000-4-2 : 2009
EN 61000-4-3 : 2006+A1 : 2008+A2 : 2010
EN 61000-4-4 : 2012
EN 61000-4-5 : 2014
EN 61000-4-6 : 2014
EN 61000-4-8 : 2010
EN 61000-4-11 : 2004

Based on the EMC Directive 2014/30/EU, Medical Device Directive 93/42/EEC and the specifications of the customer, one sample of the designated product has been tested in our laboratory and found to be in compliance with the EMC standards cited above.



TAF 0905
FCC CAB Code TW1104, TW0905
NVLAP Lab Code 200575-0
IC Code 4699A
VCCI Accep. No. R-11527, C-11609, T-11441, G-10010,
C-20010, T-20009, G-614



Central Research Technology Co.
EMC Test Laboratory
11, Lane41, Fushuen St., Jungshan Chiu,
Taipei, Taiwan, 104, R.O.C.
Tel : 886-2-25984568
Fax: 886-2-25984546

(Tsun-Yu Shih/ General Manager)

Date: October 2, 2018