

Eljintek, Inc.
Ms. Lydia Zhang
-
6F-1, No.286-3, Shin Ya Road
Chien Chen Dist.
Kaohsiung 806
Taiwan

Date : 13.11.2018
Our ref. : SCL ZTW1
Your ref.: EZ-18102018-Co.MW

Ref : TA License

Type of Equipment : Switching Power Adapter
Model Designation : See Certificate
Certificate No. : TA 50422408 0001
Report No. : 50144994 003

Dear Ms. Lydia Zhang,

Enclosed please find above certification documents.
Please forward these originals to the certificate holder.

If you contact our office, please quote our reference above.

We thank you for your cooperation.

the certificate holder is: MEAN WELL Enterprises Co., Ltd.

With kind regards,

Certification Body



Dipl.-Ing. (FH) A. Klinker

Enclosure

TÜV RHEINLAND TAIWAN LTD.

11F., No. 758, Sec. 4, Bade Rd.,
Songshan Dist., Taipei City 105,
Taiwan R. O. C.
Tel. (02) 2172-7000
Fax (02) 2172-1322
<http://www.tuv.com>

TAICHUNG BRANCH:
No. 9, Lane 36, Sec. 3, Minsheng Rd.,
Daya Dist., Taichung City 428,
Taiwan, R. O. C.
Tel. (04) 2521-8888
Fax (04) 2521-8899

KAOHSIUNG BRANCH:
27F-3, No.80, Minzu 1st Rd.,
Sanmin Dist., Kaohsiung City 807,
Taiwan, R. O. C.
Tel. (07) 262-6000
Fax (07) 380-0577

MEAN WELL Enterprises Co., Ltd.
Ms. Karen Lin Assistant Manager
Certification Center
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Date : 13.11.2018
Our ref. : SCL ZTW1
Your ref.: EZ-18102018-Co.MW

Ref : TA License

Type of Equipment : Switching Power Adapter
Model Designation : See Certificate
Certificate No. : TA 50422408 0001
Report No. : 50144994 003

Dear Ms. Karen Lin,

The above specified technical equipment has been tested and found to be in accordance with the relevant requirements.

Enclosed you will find the certificate of approval.

If cancellation of the certificate is submitted by November 15 in a given year, no fee will be charged for the following year.

We appreciate your kind support and would like to offer our assistance and continuous services in the future.

Please contact our office for approval of your new products.

With kind regards,

Certification Body



Dipl.-Ing. (FH) A. Klinker

cc: Eljintek, Inc.

Enclosure

TÜV RHEINLAND TAIWAN LTD.

11F., No. 758, Sec. 4, Bade Rd.,
Songshan Dist., Taipei City 105,
Taiwan R. O. C.
Tel. (02) 2172-7000
Fax (02) 2172-1322
<http://www.tuv.com>

TAICHUNG BRANCH:
No. 9, Lane 36, Sec. 3, Minsheng Rd.,
Daya Dist., Taichung City 428,
Taiwan, R. O. C.
Tel. (04) 2521-8888
Fax (04) 2521-8899

KAOHSIUNG BRANCH:
27F-3, No.80, Minzu 1st Rd.,
Sanmin Dist., Kaohsiung City 807,
Taiwan, R. O. C.
Tel. (07) 262-6000
Fax (07) 380-0577

Certificate



Certificate no.

TA 50422408 01

License Holder:

MEAN WELL Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

030-0000334704

Test report no.:

ZTW1-SCL 50144994 003

Client Reference: EZ-18102018-Co.MW

Tested to:

IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1+A12
IEC 60601-1-11:2015
EN 60601-1-11:2015

Certified Product: (Switching Power Adapter)

License Fee - Units

Type Designation	: GSM60Ebzzzzz (MW)	5
b stands for	: 05, 07, 09, 12, 15, 18, 24, 28, 48 or 12A	1
z stands for	: 0 to 9, A to Z, hyphen or blank	1
Rated Voltage	: AC 100-240V, 50/60Hz	
Rated Current	: 1.5-0.8A	
Output Voltage	: see Appendix	
Output Current	: see Appendix	
max. Ambient Temperature	: 40°C	
Protection Class	: II	

Remark: 1. The usability requirements of the standard were not addressed.
2. This certificate is valid for maximum 5 years.

7

Licensed Test mark:



IEC 60601-1
EN 60601-1
IEC 60601-1-11
EN 60601-1-11

www.tuv.com
ID 1419063683

TÜV Rheinland Taiwan Ltd

Signature



Dipl.-Ing. (FH) A. Klinker

Date of Issue

(day/mo/yr)

13/11/2018

Certificate



Certificate no.

TA 50422408 02

License Holder:

MEAN WELL Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

030-0000334704

Test report no.:

ZTW1-SCL 50144994 003

Client Reference: EZ-18102018-Co.MW

Tested to:

IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1+A12
IEC 60601-1-11:2015
EN 60601-1-11:2015

Certified Product: (Switching Power Adapter)

License Fee - Units

as page 01

Addition

Type Designation	: GEM60Ibzzzzz (MW)	1
b stands for	: 05, 07, 09, 12, 15, 18, 24, 28, 48 or 12A	1
z stands for	: 0 to 9, A to Z, hyphen or blank	1

This certificate is valid for maximum 5 years.

3

Licensed Test mark:



TÜV Rheinland Taiwan Ltd

Signature



Dipl.-Ing. (FH) A. Klinker

Date of Issue

(day/mo/yr)

13/11/2018

Appendix to TÜV approved Certificate No.: TA 50422408

Certified Product : **Switching Power Adapter**
Report Number : **50144994 003**
Type Designation : **GSM60Ebzzzzz, GSM60Ibzzzzz**
(b can be 05, 07, 09, 12, 15, 18, 24, 28, 48 or 12A;
z can be 0 to 9, A to Z, hyphen or blank)

Model name	Output Voltage (Vdc)	Output Current (A)	Max. output power (W)
GSM60E05zzzzz, GEM60I05zzzzz	5-6	6.00-5.00	30
GSM60E07zzzzz, GEM60I07zzzzz	6-8	6.00-5.63	45.04
GSM60E09zzzzz, GEM60I09zzzzz	8-11	5.63-5.00	55
GSM60E12Azzzzz, GEM60I12Azzzzz	11-13	5.0-4.62	60
GSM60E12zzzzz, GEM60I12zzzzz	12	4.5	54
GSM60E15zzzzz, GEM60I15zzzzz	13-16	4.62-3.75	60
GSM60E18zzzzz, GEM60I18zzzzz	16-21	3.75-2.86	60.06
GSM60E24zzzzz, GEM60I24zzzzz	21-27	2.86-2.23	60.21
GSM60E28zzzzz, GEM60I28zzzzz	27-33	2.23-1.82	60.06
GSM60x48zzzzz, GEM60I48zzzzz	33-58	1.82-1.04	60.32

Date: 13th November, 2018



Certification Body

Dipl.-Ing. (FH) A. Klinker