

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST  
CERTIFICATES FOR ELECTRICAL EQUIPMENT  
(IECEE) CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE  
CERTIFICATS D'ESSAIS DES EQUIPEMENTS  
ELECTRIQUES (IECEE) METHODE OC

CB TEST CERTIFICATE  
CERTIFICAT D'ESSAI OC

Product  
Produit

Medical Type Switching Power Supply

Name and address of the applicant  
Nom et adresse du demandeur

MEAN WELL Enterprises Co.,Ltd.  
No. 28, Wuquan 3rd Rd.,  
Wugu Dist., New Taipei City 248, Taiwan

Name and address of the manufacturer  
Nom et adresse du fabricant

MEAN WELL Enterprises Co.,Ltd.  
No. 28, Wuquan 3rd Rd.,  
Wugu Dist., New Taipei City 248, Taiwan

Name and address of the factory  
Nom et adresse de l'usine

MEAN WELL Enterprises Co.,Ltd.  
No. 28, Wuquan 3rd Rd.,  
Wugu Dist., New Taipei City 248, Taiwan

Note: When more than one factory, please report on page 2  
Note: Lorsque il y plus d'une usine, veuillez utiliser la 2<sup>ème</sup> page

Ratings and principal characteristics  
Valeurs nominales et caractéristiques principales

AC 100-240V; 50/60Hz; 1.3-0.6A; Class I;  
for output ratings refer to report no. 50062860 001  
page 7

Trade mark (if any)  
Marque de fabrique (si elle existe)

Trademark of MEAN WELL

Model/type Ref.  
Ref. de type

GSM90Ax  
(x = 12, 15, 19, 24 or 48 to denote different output ratings)

Additional information (if necessary may also be  
reported on page 2)  
Les Information complémentaire (si nécessaire,  
peuvent être indiqués sur la 2<sup>ème</sup> page)

Models differ in model name, output ratings, windings of  
main transformer T1 T1 (different numbers of secondary  
windings, gauge, turns for different output ratings) and the  
ratings of components; for the details see test report

PUBLICATION

EDITION

A sample of the product was tested and found  
to be in conformity with  
Un échantillon de ce produit a été essayé et a été  
considéré conforme à la

IEC 60601-1:2005 + A1  
IEC 60601-1-6:2010 + A1  
IEC 62366:2007 + A1  
for national deviations see test report

As shown in the Test Report Ref. No. which forms part  
of this Certificate  
Comme indiqué dans le Rapport d'essais numéro de  
référence qui constitue une partie de ce Certificat

50062860 003

This CB Test Certificate is issued by the National Certification Body  
Ce Certificat d'essai OC est établi par l'Organisme National de Certification



Date: 02.05.2017

TÜV Rheinland LGA Products GmbH  
Tillystraße 2 · 90431 Nürnberg, Germany  
Phone + 49 221 806-1371  
Fax + 49 221 806-3935  
Mail: cert-validity@de.tuv.com  
Web: www.tuv.com

Signature:

Dipl.-Ing. W. Hsu

