

Certificate



Certificate no.

TA 50341433 01

License Holder:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

Mean Well Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Test report no.: ZTW1-LinPr 10054962 001 **Client Reference:** FL-14122015-GSM90A

Tested to: IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1+A12

Certified Product: (Switching Mode Power Supply)

License Fee - Units

Type Designation	: GSM90Ax (MW)	8
x stands for	: 12, 15, 19, 24 or 48	1
Rated Voltage	: AC 100-240V, 50/60Hz	
Rated Current	: 1.3-0.6A	
Outputs	: see constructional dataform	
max. Power Output	: see constructional dataform	
max. Ambient Temperature	: 40°C	
max. Operational Altitude	: 3000m	
Protection Class	: I	

Remarks:

- 1) The risk management requirements and usability of the standard were not addressed;
- 2) This certificate is valid for maximum 5 years.

9

Appendix: 1


Licensed Test mark:



TÜV Rheinland Taiwan Ltd

Signature

W. Hsu
Dipl.-Ing. W. Hsu



Date of Issue

(day/mo/yr)

03/05/2016

Certificate



Certificate no.

TA 50341433 02

License Holder:

MEAN WELL Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

MEAN WELL Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Test report no.: ZTW1- 10054962 002

Client Reference: FL_31082016_TA mod.

Tested to:

IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1+A12

Certified Product: (Switching Mode Power Supply)

License Fee - Units

as page 01

Change

License Holder : see above

Factory : see above

This certificate is valid for maximum 5 years.

Appendix: 1


Licensed Test mark:



TÜV Rheinland Taiwan Ltd

Signature

Dipl.-Ing. W. Hsu



Date of Issue

(day/mo/yr)

17/04/2017