


IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE)  
CB SCHEME

## CB TEST CERTIFICATE

Product	AC/DC Medical Adaptor
Name and address of the applicant	MEAN WELL Enterprises Co., Ltd. No. 28, Wuquan 3rd Road, Wugu Dist., New Taipei City 248 Taiwan
Name and address of the manufacturer	MEAN WELL Enterprises Co., Ltd. No. 28, Wuquan 3rd Road, Wugu Dist., New Taipei City 248 Taiwan
Name and address of the factory <i>Note: When more than one factory, please report on page 2</i>	MEAN WELL Enterprises Co., Ltd. No. 28, Wuquan 3rd Rd., Wugu Dist., New Taipei City 248 Taiwan <input checked="" type="checkbox"/> <a href="#">Additional Information on page 2</a>
Ratings and principal characteristics	100-240Vac, 50/60Hz, 1.0-0.5A for GSM40Ax series, 1.4-0.7A for GSM60Ax series Output: Refer to Test Report for details.
Trademark (if any)	MEAN WELL 
Type of Customer's Testing Facility (CTF) Stage used	
Model / Type Ref.	GSM40Ax series, GSM60Ax series See Page 2
Additional information (if necessary may also be reported on page 2)	National Differences specified in the CB Test Report. <input checked="" type="checkbox"/> <a href="#">Additional Information on page 2</a>
A sample of the product was tested and found to be in conformity with	IEC 60601-1(ed.3)
As shown in the Test Report Ref. No. which forms part of this Certificate	180102601 issued on 2018-01-17

This CB Test Certificate is issued by the National Certification Body



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)Date: 2018-01-22  
Original Issue Date: 2013-08-09

Signature:

Jan-Erik Storgaard



Ref. Certif. No.

**DK-34275-A2-UL**

**Model Details:**

GSM40Ax series (x can be 05, 07, 09, 12, 13.5-KOD, 15, 18, 24, 28 or 48)  
GSM60Ax series (x can be 05, 07, 09, 12,15,18, 24 or 48)

**Factories:**

MEAN WELL (GUANGZHOU) ELECTRONICS CO., LTD HUADU BRANCH  
No.11 Jingu South Road, Huadong Town, Huadu District, Guangzhou  
China

SuZhou MEAN WELL Technology Co., Ltd.

No. 77, Jian-min Road, Dong-qiao, Pan-yang Ind. Park, Huang-dai Town, Xiang-cheng District, Suzhou, Jiangsu  
215152  
China

**Additional Information:**

Risk Management was not included in this investigation.

The original report was modified to include the following changes/additions:

- Add new model: GSM40A28 (output: 28 Vdc/1.42 A).
- Updated Client Declaration Table for the new model.
- Change Test item description from" Switching Adaptor" to AC/DC Medical Adaptor.
- Chang the font case of Trade Mark form "Mean Well" to "MEAN WELL"
- Update Applicant, Manufacturer and Factory information.

**Additional information (if necessary)**



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2018-01-22  
Original Issue Date: 2013-08-09

Signature:

Jan-Erik Storgaard

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST  
CERTIFICATES FOR ELECTRICAL EQUIPMENT  
(IECEE) CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE  
CERTIFICATS D'ESSAIS DES EQUIPEMENTS  
ELECTRIQUES (IECEE) METHODE OC

**CB TEST CERTIFICATE****CERTIFICAT D'ESSAI OC**

Product  
Produit

Switching Adaptor

Name and address of the applicant  
Nom et adresse du demandeur

Mean Well Enterprises Co., Ltd.  
No. 28, Wuquan 3rd Road, Wugu Dist., New Taipei City 248,  
Taiwan

Name and address of the manufacturer  
Nom et adresse du fabricant

Mean Well Enterprises Co., Ltd.  
No. 28, Wuquan 3rd Road, Wugu Dist., New Taipei City 248,  
Taiwan

Name and address of the factory  
Nom et adresse de l'usine

Mean Well Enterprises Co., Ltd.  
No. 28, Wuquan 3rd Rd., Wugu Dist., New Taipei City 248,  
Taiwan

Note: When more than one factory, please report on page 2  
Note: Lorsque il y a plus d'une usine, veuillez utiliser la 2<sup>ème</sup> page

Additional Information on page 2

Ratings and principal characteristics  
Valeurs nominales et caractéristiques principales

100-240 V ac, 50/60 Hz,  
1.0-0.5 A for GSM40Ax series  
1.4-0.7 A for GSM60Ax series  
Output: Refer to Test Report for details.

Trademark (if any)  
Marque de fabrique (si elle existe)

Mean Well



Type of Manufacturer's Testing Laboratories used  
Type de programme du laboratoire d'essais  
constructeur

GSM40Ax series, GSM60Ax series  
See Page 2

Model / Type Ref.  
Ref. De type

Additional Information on page 2

Additional information (if necessary may also be  
reported on page 2)  
Les informations complémentaires (si nécessaire,,  
peuvent être indiqués sur la 2<sup>ème</sup> page

A sample of the product was tested and found  
to be in conformity with  
Un échantillon de ce produit a été essayé et a été  
considéré conforme à la

IEC 60601-1-2  
A

As shown in the Test Report Ref. No. which forms  
part of this Certificate  
Comme indiqué dans le Rapport d'essais numéro de  
référence qui constitue partie de ce Certificat

161001402 issued on 2016-11-07

This CB Test Certificate is issued by the National Certification Body

Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2016-11-09

Signature:

Original Issue Date: 2013-08-09

Jan-Erik Storgaard



Model Details:

GSM40Ax series

GSM60Ax series

Á

Factories:

MEAN WELL (GUANGZHOU) ELECTRONICS CO., LTD HUADU BRANCH
No.11 Jingu South Road, Huadong Town, Huadu District, Guangzhou,
China

SuZhou Mean Well Technology Co., Ltd.

No. 77, Jian-min Road, Dong-qiao, Pan-yang Ind. Park, Huang-dai Town, Xiang-cheng District, Suzhou, Jiangsu
215152,
China

Additional Information:

The original report was modified to include the following changes/additions:

- Added new model with new label.
- Updated Client Declaration Table for the new model.
- Updated Device type from sub-assembly to component.
- Updated factory information and manufacturer's declaration letter.

National Differences specified in the CB Test Report.

Risk Management was not included in this investigation.

Additional information (if necessary)

Information complémentaire (si nécessaire)



- UL (US), 333 Pflingsten Rd IL 60062, Northbrook, USA
UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2016-11-09

Original Issue Date: 2013-08-09

Signature: [Handwritten signature]

Jan-Erik Storgaard

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST  
CERTIFICATES FOR ELECTRICAL EQUIPMENT  
(IECEE) CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE  
CERTIFICATS D'ESSAIS DES EQUIPEMENTS  
ELECTRIQUES (IECEE) METHODE OC

**CB TEST CERTIFICATE****CERTIFICAT D'ESSAI OC**

Product  
Produit

Switching Adaptor

Name and address of the applicant  
Nom et adresse du demandeur

Mean Well Enterprises Co., Ltd.  
No. 28, Wuquan 3rd Road, Wugu Dist., New Taipei City 248,  
Taiwan

Name and address of the manufacturer  
Nom et adresse du fabricant

Mean Well Enterprises Co., Ltd.  
No. 28, Wuquan 3rd Road, Wugu Dist., New Taipei City 248,  
Taiwan

Name and address of the factory  
Nom et adresse de l'usine

Mean Well Enterprises Co., Ltd.  
No. 28, Wuquan 3rd Rd., Wugu Dist., New Taipei City 248  
Taiwan

Note: When more than one factory, please report on page 2  
Note: Lorsque il y a plus d'une usine, veuillez utiliser la 2<sup>ème</sup> page

Additional Information on page 2

Ratings and principal characteristics  
Valeurs nominales et caractéristiques principales

100-240 V ac, 50/60 Hz,  
1.0-0.5 A for GSM40Ax series  
1.4-0.7 A for GSM60Ax series  
Output: Refer to Test Report

Trademark (if any)  
Marque de fabrique (si elle existe)

Mean Well or



Type of Manufacturer's Testing Laboratories used  
Type de programme du laboratoire d'essais  
constructeur

GSM40Ax series, GSM60Ax series  
See Page 2

Model / Type Ref.  
Ref. De type

Additional Information on page 2

Additional information (if necessary may also be  
reported on page 2)  
Les informations complémentaires (si nécessaire,,  
peuvent être indiqués sur la 2<sup>ème</sup> page

A sample of the product was tested and found  
to be in conformity with  
Un échantillon de ce produit a été essayé et a été  
considéré conforme à la

IEC 60601-1(ed.3)

As shown in the Test Report Ref. No. which forms  
part of this Certificate  
Comme indiqué dans le Rapport d'essais numéro de  
référence qui constitue partie de ce Certificat

1302034002 issued on 2013-08-06

This CB Test Certificate is issued by the National Certification Body

Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2013-08-09

Signature:

Jan-Erik Storgaard



Ref. Certif. No.

**DK-34275-UL**

**Model Details:**

GSM40Ax series, GSM60Ax series (x can be 05, 07, 09, 12, 15, 18, 24 or 48)

**Factories:**

Mean Well (GUANGZHOU) Electronics Co., Ltd.  
2nd Floor, No. A Building, Yuean Ind. Park, Dongpu Town, TianHe Dist, Guangzhou  
China

SuZhou Mean Well Technology Co., Ltd.  
No. 77, Jian-min Road, Dong-qiao, Pan-yang Ind. Park, Huang-dai Town, Xiang-cheng District, Suzhou, Jiangsu  
215152  
China

**Additional Information:**

Additionally evaluated to EN 60601-1: 2006.  
National Differences specified in the CB Test Report.  
Risk Management was not included in this investigation.

**Additional information (if necessary)**

**Information complémentaire (si nécessaire)**



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2013-08-09

Signature:

Jan-Erik Storgaard