



中国认可
国际互认
检测
TESTING
CNAS L1842



佛山中纺联检验技术服务有限公司
CNTAC Testing Service Co.,Ltd.(Foshan)

检测报告(Testing Report)	防伪查询网址(Security website): www.fcl-sz.org.cn 防伪码(Security code): 7818853801
报告编号(Report No): ZFLJ2615386	Page 1 of 4



Applicant Information (客户信息)

Applicant Name(委托单位) : Dongguan Zhanhong Weaving String Co., Ltd.(东莞展宏织带有限公司)
Applicant Address(委托单位地址) : Xitailong Industrail Area, Pearl Road, Shatian Town, Dongguan City, Guangdong Province, China.(广东省东莞市沙田镇西太隆明珠路 44 号)
Manufacturer(生产单位) : Dongguan Zhanhong Weaving String Co., Ltd.(东莞展宏织带有限公司)

Sample Information (样品信息)

Sample Description(样品描述) : Disposable 3PLY masks (普通平面防护口罩)
Fiber Components(纤维成分) : Non-woven fabric,melt-blown fabric (无纺布,熔喷布)
Sample Quantity(样品数量) : 50 pieces (50 个)
Colour(颜色) : White (白色)/ Blue (蓝色)

- Sample Receiving date (收到样品的日期) : 2020-04-20
- Report date (报告日期) : 2020-05-03
- The original sample is stucked on the last paper. (送检样品原样粘贴在本报告的末页)

Test Performed (检验标准)

Judgement according to(评定依据):
GB/T 32610-2016 Technical Specification of daily protective mask
日常防护型口罩技术规范
- Selected test(s) as requested by applicants. (本实验室根据客户要求完成以下检测内容)

Pronounce (声明)

The results shown in this report refer only to sample(s) tested unless otherwise stated.All the items are performed in the standard conditions, except the noted cases.(除非特别说明, 测试结果只对来样负责。所有测试项目均在标准规定的环境下进行, 有注明除外。)
Except for the requirement of the client, the test results and the conformity judgement of this report do not take the uncertainty of the test results into account.(除非客户要求, 本报告检测结果及符合性判定不考虑测量结果的不确定度。)

Signed for and on behalf of
CNTAC Testing Service Co.,Ltd.(Foshan)

Approved by
(批准)

(Except for microbiological project 微生物项目除外)

(Microbiological project 微生物项目)





中国认可
国际互认
检测
TESTING
CNAS L1842



检测报告(Testing Report)	防伪查询网址(Security website): www.fcl-sz.org.cn
	防伪码(Security code): 7818853801
报告编号(Report No): ZFLJ2615386	Page 2 of 4



--- Test Result (测试结果) ---

Test Result(测试结果)	Requirements(评定条件)	Judgement(判定)	
1. Filtration Efficiency(Salt medium) (过滤效率(盐性介质)) <u>GB/T 32610-2016 Appendix A(附录 A)</u>			
Air flow: 85L/min, Aerosol: NaCl, (气体流量 85L/min, 气溶胶颗粒: NaCl)			
Unit(单位): <%>			
Minimum value(最小值)	95.0	Level II (II级) ≥ 95	Pass(合格)
2. Tensile Strength of the mask belt and the connection between the mask belt and the mask body (口罩带及口罩带与口罩体的连接处断裂强力) <u>GB/T 32610-2016 Section 6.9(第 6.9 条款)</u>			
Unit(单位): <N>			
Result(测试结果)	24	≥ 20	Pass(合格)
3. pH Value(pH 值) <u>GB/T 7573-2009</u>			
Using KCl solution as the extracting solution.(萃取介质: 氯化钾溶液)			
Result(测试结果)	6.0	4.0~8.5	Pass(合格)
4. Exhalation Resistance(呼气阻力) <u>GB/T 32610-2016 Section 6.8(条款 6.8)^[1]</u>			
Air flow: 85L/min(通气量 85 L/min)			
Unit(单位): <Pa>			
Maximum value(最大值)	54.8	≤ 145	Pass(合格)
5. Inhalation Resistance(吸气阻力) <u>GB/T 32610-2016 Section 6.7(条款 6.7)^[1]</u>			
Air flow: 85L/min(通气量 85 L/min)			
Unit(单位): <Pa>			
Maximum value(最大值)	65.0	≤ 175	Pass(合格)
6. Microbiological(微生物) <u>GB 15979-2002 Appendix B(附录 B)</u>			
Unit(单位): <CFU/g>			
Total number of bacterial colonies (细菌菌落总数)	<20	Total number of bacterial colonies(细菌菌落总数) ≤ 200	Pass(合格)
Pathogenic purulent bacteria(致病性化脓菌)		Pathogenic purulent bacteria	
-Pseudomonas aeruginosa Not Detectable (绿脓杆菌)	未检出	Not Detectable (致病性化脓菌) 不得检出	
-Staphylococcus aureus Not Detectable (金黄色葡萄球菌)	未检出		
-Hemolytic streptococcus Not Detectable (溶血性链球菌)	未检出		
The total number of fungal colonies (真菌菌落总数)	<20	The total number of fungal colonies(真菌菌落总数) ≤ 100	
Coli group Not Detectable (大肠菌群)	未检出	Coli group Not Detectable (大肠菌群) 不得检出	





中国认可
国际互认
检测
TESTING
CNAS L1842



佛山中纺联检验技术服务有限公司
CNTAC Testing Service Co.,Ltd.(Foshan)

检测报告(Testing Report)	防伪查询网址(Security website): www.fcl-sz.org.cn 防伪码(Security code): 7818853801
	报告编号(Report No): ZFLJ2615386



Test Result(测试结果)

Requirements(评定条件)

Judgement(判定)

7. Appearance Inspection(外观要求) GB/T 32610-2016 Section 6.1(条款 6.1)

Result(测试结果)	Accordance (符合)	Mask surface should not be damaged, oil stains, deformation and other obvious defects(口罩表面不应有破损、油污斑渍、变形及其他明显的缺陷)	Pass(合格)
--------------	-----------------	--	----------

Remark(备注): ^[1]=This method is beyond the scope of CNAS accreditation .
(此方法不在本实验室 CNAS 认可能力范围内。)





中国认可
国际互认
检测
TESTING
CNAS L1842



佛山中纺联检验技术服务有限公司
CNTAC Testing Service Co.,Ltd.(Foshan)

检测报告(Testing Report)	防伪查询网址(Security website): www.fcl-sz.org.cn 防伪码(Security code): 7818853801
报告编号(Report No): ZFLJ2615386	Page 4 of 4



Original (送检样品原样)



===== End of Report (检测报告结束) =====

