

Pregnancy Test Midstream

It is a tool used by women to detect pregnancy. The principle of the test is to detect HCG in human urine.

Description :

Human chorionic gonadotropin (HCG) quantitative test kit (colloidal gold) adopts highly specific antigen-antibody reaction and colloidal gold immune chromatography to detect HCG. In a specific wavelength, the amount of colloidal gold particles is associated with the absorption and scattering to light. By using the photoelectric sensor to detect the absorption and scattering of colloidal gold particles to the light, the amount of antigen to be determined can be obtained.

Specification: midstream

Technical Characteristics

1. Meeting the "human chorionic gonadotropin (HCG) quantitative test kit (colloidal gold) registered product standard" requirement.
2. The accordance rate of negative reference material: the concentration should be fully detected as 0 mIU/ml after the determination of 20 negative serum samples.
3. Accuracy: the HCG serum with the specified value is used for the determination. Each of the concentration is taken its mean for five times and the mean deviation of its respective concentration is not more than 10%.
4. Linear correlation coefficient: $r \geq 0.99$
5. Minimum detectable amount: the concentration is not higher than 25 mIU/ml.
6. Sensitivity: the measured concentration of reagent whiteboard is 0 mIU/ml.
7. Precision: the repeatable requirement of intra- and inter-regent is not more than 10%.